



Summit Technology 2717 N. Main St. Suite 15, Walnut Creek, CA 94597  
1-925-944-1212 fax:1-925-944-7126 www.powersight.com

\*\*\*\* Calibration Certificate \*\*\*\*\* Calibration Certificate \*\*\*\*\* Calibration Certificate \*\*\*\*\*

Date: May 9, 2012

Certification # **C1205091**

Cert. for:

Summit Technology

2717 North Main St

Suite 15

Walnut Creek, CA 94597

Phone: 925-944-1212

Fax: 925-944-7126

Email: Support@PowerSight.com

The following information is in regards to the calibration of the following equipment:

**Serial number 04500**

**Our records indicate that it was calibrated on May 9, 2012**

**Calibration was performed using NIST traceable equipment.**

**Specifically, the following test equipment was employed during calibration:**

Valhalla Scientific 2703 AC Voltage Standard

Valhalla Scientific 2555A AC/DC Current Calibrator

Hewlet Packard 34401A Digital 6-1/2 Digit Multi-Meter

See attached sheet for calibration measurement data.

**Summit Technology recommends calibration of meters and probes every 12 months.**

If you have any questions, contact support@powersight.com.

Best regards,

Calibration Technician